

## Relative Humidity Sensor 3445 is designed for use with the Aanderaa Automatic Weather Station 2700, Dataloggers 3660 or 3634 and other sensor read-out units.

Automatic Weather Station AWS 2700

Relative Humidity is one of the basic parameters in meteorology and a Relative Humidity (RH) Sensor will normally be included in a weather station. Sensor 3445 is designed to fit Aanderaa Automatic Weather Station 2700, but can also be used as a separate unit.

The 3445 sensor consists of the RH Probe 3447 and a Radiation Screen 3446. The probe has a sensing element at one end, a 6-pin receptacle at the other, and an electronic board in between, all of which are molded in durotong foam into a cylindrically shaped probe.

The probe is furnished with a standard Aanderaa sensor foot to which the screen is fastened. The


screen protects the probe against direct sunshine and rain. The RH Sensor 3445 is based upon an integrated circuit, IH-3602-L from Hy-Cal Engineering, as the sensing element. The circuit employs the well proven capacitive polymer to sense humidity.

The sensor also incorporates a millipore filter which reduces the atmospheric pollution of the sensing element to a minimum.

The circuit has a linear output voltage of between 0 – 100% relative humidity. A built-in AD converter and a micro-controller changes this voltage to the standard Aanderaa SR10 digital output signal.

## **Specifications**

**RELATIVE HUMIDITY SENSOR 3445** 



PIN CONFIGURATION

Receptacle, exterior view; bushing =o; pin = •

Not connected Bridge voltage Control Voltage Output signal System ground Not connected

Accuracy: ±2% RH Resolution: 0.1% IH-3206-L Sensing Element: Sensor Output: Digital SR10 Operating Temperature: -40 to +65°C 1.5 min Time Constant:

Current Consumption: Maximum 1.2mA when

operating

Electrical Connection: Automatic Weather

Station 2700's Sensor Cross Arm (Watertight Plug 2828)

0 to 100% relative humidity (RH)

Material and Finish:

Range:

Probe: Durotong foam and titanium

Radiation Screen: White nylon

Dimensions: Radiation Screen, Ø 60x 128 mm

140g Net Weight: 500g Gross Weight:

Packing: Cardboard box,

> 300x200x210mm Two years against faulty materials and workmanship

Note! Check the millipore filter for contamination from time to time depending on the grade of pollution in the area and replace if necessary.

Connecting Cable 2842 (10 meter cable with watertight plugs), is available for connecting this sensor to the Aanderaa Datalogging systems. Other lengths, or separate plugs and cables are available on request.

## **Calibration**

The sensor is calibrated at 0% and 99% RH. 0% RH is obtained in a vacuum chamber and 99%RH in a chamber with moisture saturated air. The sensor is connected to a Datalogger 3660, giving the following raw data readings (N): The sensor readings are also compared with a psychrometer, Lambrecht Type 00.07610.000012. No. 670773.001.

% RH	0	99
Reading, N		

## Serial No:

Warranty:

These readings give a set of coefficients used when converting a sensor reading to engineering units. The coefficients are:

А	С	0
В	D	0

To convert raw data to engineering units use the formula:  $%RH = A + BN + CN^2 + DN^3$ 

Date: Sign:



Visit our Web site for the latest version of this document and more information

www.aanderaa.com

Aanderaa is a trademark of Xylem Inc. or one of its subsidiaries. © 2012 Xylem, Inc. D271 November 2012

Aanderaa Data Instruments AS Sanddalsringen 5b, Postboks 103 Midtun, 5828 Bergen, Norway Tel +47 55 60 48 00 Fax +47 55 60 48 01